

## REQUIREMENT SPECIFICATIONS

### REQUEST FOR PROPOSAL FOR THE SUPPLY AND DELIVERY OF BLOOD TUBING SET AND ARTERIAL VENOUS FISTULA NEEDLES

#### 1. Introduction

- 1.1 The National Kidney Foundation (“**NKF**”) wishes to appoint a contractor (the “**Contractor**”) for the supply and delivery of Blood Tubing Set And Arterial Venous Fistula Needles as described in these Requirement Specifications (the “**Goods**”), to all its dialysis centres in Singapore.

#### 2. Product Specification

- 2.1 The Goods shall conform to the detail specification in **Annex A**.

#### 3. Contract Duration, Quantity Requirements and Delivery Schedule

- 3.1 The Contractor shall supply the Goods over a period of **thirty-six (36) months from 1 January 2027 to 31 December 2029**.
- 3.2 NKF may extend the Contract Period, on the same terms as those contained herein (including on the same Contract Price), for a further duration not exceeding 24 months, by serving written notice of the Company’s desire to renew the term hereof, which notice shall be given not less than 30 days prior to the expiry of the initial Contract Period.
- 3.3 NKF’s estimated consumption for the Goods over a period of thirty-six (36) months is set out as below:

| Sn | Item Description                                 | Estimated Quantity for 36 months |
|----|--|----------------------------------|
| 1  | Blood Tubing Set                                 | 2,411,940 pcs                    |
| 2  | Arterial Venous Fistula Needles, Size <u>14G</u> | 219,900 pcs                      |
| 3  | Arterial Venous Fistula Needles, Size <u>15G</u> | 2,397,650 pcs                    |
| 4  | Arterial Venous Fistula Needles, Size <u>16G</u> | 1,069,720 pcs                    |
| 5  | Arterial Venous Fistula Needles, Size <u>17G</u> | 37,290 pcs                       |

- 3.3 The Goods shall be delivered monthly to such of NKF’s dialysis centres as NKF shall stipulate from time to time. Please refer to **Annex B** for the detail list of NKF’s Dialysis Centres (as at date of this RFP). For the avoidance of doubt, NKF reserves the right at any time to increase or decrease the number of and to vary and /or change the location of any or all of the listed Dialysis Centres. Avoid delivery from: 1300hours to 1400hours

Authorised Signature: \_\_\_\_\_ Vendor’s stamp : \_\_\_\_\_

**4. Submission of RFP bids**

- 4.1 Each Vendor should provide the price quote in the Price Schedule.
- 4.2 All quotations submitted by the Vendor must indicate the prices applicable for the estimated numbers of Goods specified in paragraph 3.2 above.
- 4.3 The Vendor, at the point of submission of its RFP bids, is required to supply at its own costs the required samples as provided in point 5 for each offered product stipulated in its RFP bids. The NKF shall be at liberty to call for further submission of such samples as required until the samples submitted are in accordance with the requirements of the RFP. Samples after approval shall indicate the standards to be maintained for the duration of the Contract.
- 4.4 Each vendor is required to indicate the following information for the proposed products in the Price Schedule:
- 4.4.1 The minimum delivery lead time
  - 4.4.2 Country of manufacturer
  - 4.4.3 Location of storage, including alternative site

**5. Submission of samples**

- 5.1 Each vendor shall submit **1 Box** of samples for each of the item in the full range of products set out in the proposal requirement in respect of which they are quoting, free-of charge and non-returnable, to the Company at the following address:

**National Kidney Foundation**

81 Kim Keat Road,  
Singapore 328836  
Attention: Ms Mavis Tan  
Department: Purchasing

- 5.2 Each set of samples must be labelled individually with the following information:
- RFP Number
  - Item S/N (as per price schedule S/No)
  - Company Name
  - Product Code Number
- 5.3 The samples are to be submitted **within four (4) weeks from the Request For Proposal opening date**, or such extended time period as the Company may agree to in writing upon request by the Company, **after the submission of the Request For Proposal**. The vendor may be rejected if samples are not submitted on time.
- 5.4 If more sample are required, as may be requested by NKF from time to time, on the terms and conditions set out in the Request For Proposal.

Authorised Signature: \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_

## REQUIREMENT SPECIFICATIONS

| SPECIFICATIONS OF BLOOD TUBING SET   |   |  |                       |
|--|---|--|-----------------------|
| <b>Description of the Product:</b> Blood tubing set is required as integral part of dialysis therapy along with the Dialyzer. Blood tubing set is used for drawing blood and returning to body through Dialyzer during dialysis therapy. |   |  |                       |
| S.N.   | Parameter   | Specification  | Comply (Yes/ No/ N.A) |
| <b>1. General Specification</b>  |   |  |                       |
| 1.1  | Universal Tubing set                                  | Set should be appropriate for Fresenius, Gambro and B Braun models in all series of haemodialysis machines.          |                       |
| 1.2  | Tube material   | Biocompatible / Medical Grade PVC Tube.  |                       |
| 1.3  | Clamp   | Provide clamps on all insert ports to secure blocking performance.   |                       |
| 1.4  | Sterilization   | ETO/Steam, Radiation E-Beam/Gamma - please specify the method  |                       |
| 1.5  | Universal tubing set shall comprise                   | (1) An arterial bloodline set.<br>(2) A venous bloodline set.<br>(Both lines shall be packed in a sterile pouch.)    |                       |
| <b>2. Operating Specification</b>  |   |  |                       |
| 2.1  | Size of Blood Pump Segment                            | Inner Diameter : 8mm and Outer diameter : 12mm   |                       |
| 2.2  | Size of Arterial and Venous Blood Line                | Inner Diameter : 4.6mm and Outer diameter : 6.8mm  |                       |
| 2.3  | Length of Arterial and Venous Blood Line              | Arterial line: 3700mm to 3900mm<br>Venous line: 2500mm to 2900mm   |                       |
| 2.4  | Length of Arterial and Venous Chamber                 | Arterial Chamber: 125mm<br>Venous Chamber :112mm   |                       |
| 2.5  | Size of Inner diameter of Arterial and Venous Chamber | Arterial Chamber one end:15.7mm, other end: 16.6mm<br>Venous Chamber: 18.80mm  |                       |
| 2.6  | Size of Outer diameter of Arterial and Venous Chamber | Arterial Chamber: 20mm<br>Venous Chamber: 22mm<br>Venous Chamber: 22mm with soft material                            |                       |
| 2.7  | Pre pump Arterial pressure monitoring blood line      | Pre pump arterial pressure monitoring line 500mm in length with clamp and transducer moulded to the monitoring line. |                       |
| 2.8  | "T" junction saline infusion line.                    | It should have a clamp and a recirculation tube attached at the end.   |                       |
| 2.9  | Colour coded ports and connectors                     | a. All Ports and Connectors of Arterial line must be in RED colour.  |                       |
|  |   | b. All Ports and Connectors of Venous line must be in BLUE colour.   |                       |

Authorised Signature: \_\_\_\_\_ Vendor's stamp : \_\_\_\_\_

| S.N.                           | Parameter   | Specification  | Comply<br>(Yes/ No/ N.A) |
|--------------------------------|---|--|--------------------------|
| 2.10                           | Injection Port  | a. All injection port should have minimum width of 26mm.   |                          |
|                                |   | b. All injection port must be colour coded in RED for Arterial line and BLUE for Venous line.  |                          |
| 2.11                           | Luer Lock   | a. Arterial line must have a recirculation tube attached to the needle end.  |                          |
|                                |   | b. Both Arterial and Venous line must have luer lock connection to attach vascular access and dialyser.  |                          |
|                                |   | c. All injection ports must have bungs attached.   |                          |
| 2.12                           | Particulate Filter  | a. Venous chamber must have particulate filter, with one insertion line, with clamp attached and venous pressure monitoring line 350mm with clamp and transducer moulded to the monitoring line. |                          |
|                                |   | b. Arterial bubble chamber with insertion line with clamp attached.  |                          |
| 3. Packaging Requirement       |   |  |                          |
| 3.1                            | Each packaging must have clear label indicating the batch number or lot number, manufacturing date & expiry date, sterilisation method and priming volume.  |  |                          |
| 4. Certification Requirement   |   |  |                          |
| 4.1                            | Is product HSA registered?<br><br>Vendors to provide HSA registration and classification / attach screenshot from Singapore Medical Device Information Communication System (MEDICS) for registered and exempted products |  |                          |
| 4.2                            | Vendors to indicate and attach the quality standards and regulatory approval for the product  |  |                          |
| 4.3                            | Manufacturing license with HSA  |  |                          |
| 4.4                            | Any other relevant certification or clinical papers necessary for the product   |  |                          |
| 5. Brochures/ Catalogues/ MSDS |   |  |                          |
| 5.1                            | Vendor shall complete and provide all necessary information including product literature, brochures, catalogues or MSDS related to the products.  |  |                          |

Authorised Signature: \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_

| <b>6. Submission of Samples</b> |   | <b>Comply<br/>(Yes/ No/ N.A)</b> |
|---------------------------------|---|----------------------------------|
| 6.1                             | <p>Vendor shall submit <b>1 Box</b> of the samples for each of the items in the full range of products set out in the requirement in respect of which vendor is quoting, free-of-charge and non-returnable, to the following address:</p> <p style="text-align: center;">National Kidney Foundation<br/>81 Kim Keat Road,<br/>Singapore 328836</p> <p>Samples must come in proper packaging and artwork.</p> <p>Each sample must be labelled individually with the following information and submitted not later than <b>15 January, Thursday</b>.</p> <ul style="list-style-type: none"> <li>• Ref No: 20251202</li> <li>• Item S/N</li> <li>• Company Name</li> </ul> |                                  |

Authorised Signature: \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_

## REQUIREMENT SPECIFICATIONS

| SPECIFICATIONS OF ARTERIAL VENOUS FISTULA NEEDLES  |   |  |                          |
|--|---|--|--------------------------|
| <b>Description of the Product:</b> Fistula needle set is required as integral part of dialysis therapy along with dialyser and blood tubing set (extra-corporeal circuit). It is used for drawing and returning blood to the body through extra-corporeal circuit during dialysis therapy. |   |  |                          |
| S.N.   | Parameter   | Specification  | Comply<br>(Yes/ No/ N.A) |
| 1. General Specification   |   |  |                          |
| 1.1  | Gauge   | 14G, 15G, 16G, 17G   |                          |
| 1.2  | Needle Length   | 1 inch (25mm) with Back-Eye  |                          |
| 1.3  | Cannula   | Extra thin wall with clear kink resistant tube fitted with female luer lock connection and pinch clamp to secure blocking performance. |                          |
| 1.4  | Tube Material   | Biocompatible / Medical Grade PVC Tube   |                          |
| 1.5  | Tube length   | 300 mm   |                          |
| 1.6  | Tube volume   | 2.8 mls  |                          |
| 1.7  | Wing  | Fixed wing   |                          |
| 1.8  | Wing Colour   | The colour of the Wing should be able to clearly differentiate between the 4 sizes needles (14G,15G, 16G, 17G)                         |                          |
| 1.9  | Sterilization   | ETO/Steam, Radiation E-Beam/Gamma - please specify the method  |                          |
| 2. Packaging Requirement   |   |  |                          |
| 2.1  | Each packaging must have clear label indicating the batch number or lot number, manufacturing date & expiry date and sterilisation method.  |  |                          |
| 3. Certification Requirement   |   |  |                          |
| 3.1  | Is product HSA registered?<br><br>Vendors to provide HSA registration and classification / attach screenshot from Singapore Medical Device Information Communication System (MEDICS) for registered and exempted products |  |                          |
| 3.2  | Vendors to indicate and attach the quality standards and regulatory approval for the product  |  |                          |
| 3.3  | Manufacturing license with HSA  |  |                          |
| 3.4  | Any other relevant certification or clinical papers necessary for the product   |  |                          |
| 4. Brochures/ Catalogues/ MSDS   |   |  |                          |
| 4.1  | Vendor shall complete and provide all necessary information including product literature, brochures, catalogues or MSDS related to the products   |  |                          |

Authorised Signature: \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_

| <b>5. Submission of Samples</b> |   | <b>Comply<br/>(Yes/ No/ N.A)</b> |
|---------------------------------|---|----------------------------------|
| 5.1                             | <p>Vendor shall submit <b>1 Box</b> of the samples for each of the items in the full range of products set out in the requirement in respect of which vendor is quoting, free-of-charge and non-returnable, to the following address:</p> <p style="text-align: center;">National Kidney Foundation<br/>81 Kim Keat Road,<br/>Singapore 328836</p> <p>Samples must come in proper packaging and artwork.</p> <p>Each sample must be labelled individually with the following information and submitted not later than <b>15 January, Thursday</b>.</p> <ul style="list-style-type: none"> <li>• Ref No: 20251202</li> <li>• Item S/N</li> <li>• Company Name</li> </ul> |                                  |

Authorised Signature: \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_

**Annex B**

**LIST OF NKF DIALYSIS CENTRES**

| <b>NKF DC ADDRESS LIST</b> |   |
|----------------------------|---|
| <b>ADT</b>                 | <b>761 WOODLANDS</b><br>The Hour Glass-NKF Dialysis Centre (Admiralty Branch)<br>BLK 761 Woodlands Ave 6. #01-108. S (730761)<br>Tel : 6362 2153 / Fax : 6362 2360                      |
| <b>ALJ</b>                 | <b>102 ALJUNIED</b><br>NKF Dialysis Centre Supported by Hong Leong Foundation<br>BLK 102 Aljunied Crescent. #01-265. S (380102)<br>Tel : 6743 3572 / Fax : 6743 0817                    |
| <b>AM2</b>                 | <b>633 ANG MO KIO</b><br>NKF Dialysis Centre<br>BLK 633 Ang Mo Kio Ave 6. #01-5155. S (560633)<br>Tel : 6459 0113 / Fax : 6552 1697   |
| <b>AM3</b>                 | <b>565 ANG MO KIO</b><br>Pei Hwa Foundation-NKF Dialysis Centre<br>BLK 565 Ang Mo Kio Ave 3. #01-3401. S (560565)<br>Tel : 6552 6569 / Fax : 6552 6539                                  |
| <b>BB2</b>                 | <b>113A BUKIT BATOK</b><br>NKF Dialysis Centre Supported by Singapore Buddhist Youth Mission<br>BLK 113A Bt Batok West Ave 6. #01-01. S (651113)<br>Tel : 6513 6873 / Fax : 6322 7715   |
| <b>BED</b>                 | <b>27 NEW UPPER CHANGI</b><br>NKF Dialysis Centre Supported by San Wang Wu Ti Religious Society<br>BLK 27 New Upper Changi Rd. #01-694 S (462027)<br>Tel : 6444 4278 / Fax : 6444 4978  |
| <b>BD2</b>                 | <b>105 BEDOK</b><br>NKF Dialysis Centre Supported by Man Fatt Lam Buddhist Temple<br>BLK 105 Bedok North Ave 4. #01-2168. S (460105)<br>Tel : 6243 7349 / Fax : 6214 0210               |
| <b>BDR</b>                 | <b>213 BIDADARI</b><br>NKF Dialysis Centre Supported by Lorong Koo Chye Sheng Hong Temple<br>BLK 213 Bidadari Park Drive. #03-14 S (360213)<br>Tel : 6550 4459 / Fax : 6550 4462        |
| <b>BM2</b>                 | <b>128 BUKIT MERAH</b><br>NKF Dialysis Centre Supported by The Singapore Buddhist Lodge<br>BLK 128 Bukit Merah View. #01-22 S (150128)<br>Tel : 6878 0552 / Fax : 6878 0021             |
| <b>BPJ</b>                 | <b>274 BANGKIT</b><br>NKF Dialysis Centre Supported by New Creation Church<br>BLK 274 Bangkit Rd. #01-54. S (670274)<br>Tel : 6764 6400 / Fax : 6764 2004                               |
| <b>BP2</b>                 | <b>275 BANGKIT</b><br>NKF Dialysis Centre Supported by Le Champ (S.E.A.)<br>BLK 275 Bangkit Rd. #01-96. S (670275)<br>Tel : 6891 2782 / Fax : 6891 2592                                 |
| <b>CLE</b>                 | <b>326 CLEMENTI</b><br>Lew Foundation-NKF Dialysis Centre<br>BLK 326 Clementi Ave 5. #01-175. S (120326)<br>Tel : 6775 0668 / Fax : 6775 0891   |
| <b>CP1</b>                 | <b>IRC LEVEL 1</b><br>NKF IRC Supported by The Sirivadhanabhakdi Foundation<br>500 Corporation Road Level 1 S (649808)<br>Tel : 6359 3610 / Fax : 6251 4174                             |
| <b>CP2</b>                 | <b>IRC LEVEL 2</b><br>NKF IRC Supported by The Sirivadhanabhakdi Foundation<br>500 Corporation Road Level 2 S (649808)<br>Tel : 6359 3620 / Fax : 6251 4175                             |
| <b>FNV</b>                 | <b>465 FERNVALE</b><br>NKF Dialysis Centre Supported by Christopher Heng Tiaw Boon Memorial Fund<br>BLK 465 Fernvale Road. #01-07. S (760840)<br>Tel : 6550 4439 / Fax : 6550 4457      |
| <b>GMH</b>                 | <b>1 GHIM MOH</b><br>Woh Hup-NKF Dialysis Centre<br>1 Ghim Moh Road, #01-358. S (270001)<br>Tel : 64679200 / Fax : 64679231   |
| <b>MSD</b>                 | <b>204 MARSILING</b><br>Jo & Gerry Essery-NKF Dialysis Centre<br>BLK 204 Marsiling Drive. #01-188. S (730204)<br>Tel : 6368 0291 / Fax : 6368 0267                                      |
| <b>PNG</b>                 | <b>PUNGGOL OASIS TERRACE</b><br>NKF Dialysis Centre Supported by Ngiam Kia Hum & Family<br>BLK 681 Punggol Drive. #02-02. S (820681).<br>Tel : 62437020 / Fax : 62437022                |
| <b>PN2</b>                 | <b>ONE PUNGGOL</b><br>NKF Dialysis Centre Supported by Kwan Im Thong Hood Cho Temple<br>1 Punggol Drive. #04-06. S (828629)<br>Tel : 6970 8130 / Fax : 6970 8134                        |
| <b>PSR</b>                 | <b>180 PASIR RIS</b><br>Tampines Chinese Temple-NKF Dialysis Centre<br>BLK 180 Pasir Ris St 11. #01-06. S (510180)<br>Tel : 6583 9500 / Fax : 6583 0779                                 |
| <b>PR2</b>                 | <b>427 PASIR RIS</b><br>NKF Dialysis Centre Supported by TL Whang Foundation<br>BLK 427 Pasir Ris Drive 6. #01-35/43. S (510427)<br>Tel : 6243 1008 / Fax : 6243 4332                   |
| <b>QT1</b>                 | <b>55 STRATHMORE</b><br>NKF Dialysis Centre Supported by San Wang Wu Ti Religious Society<br>BLK 55 Strathmore Ave. #01-145. S (140055)<br>Tel : 6778 0330 / Fax : 6777 6155            |
| <b>SKG</b>                 | <b>SENGKANG COMMUNITY HOSPITAL</b><br>NKF Dialysis Centre Supported by San Wang Wu Ti Religious Society<br>BLK 1 Anchorvale St. #01-26. S (544835)<br>Tel : 6908 4501 / Fax : 6908 4504 |
| <b>SMI</b>                 | <b>101 SIMEI</b><br>NKF Dialysis Centre Supported by Kwan Im Thong Hood Cho Temple<br>BLK 101 Simei St 1. #01-892. S (520101)<br>Tel : 6785 9882 / Fax : 6786 6268                      |
| <b>SRG</b>                 | <b>201 SERANGOON</b><br>NKF Dialysis Centre Supported by IFPAS<br>BLK 201 Serangoon Central. #01-30. S (550201)<br>Tel : 6285 4113 / Fax : 6284 2553                                    |
| <b>TM1</b>                 | <b>935 TAMPINES</b><br>National Trades Union Congress/Singapore Pools-NKF Dialysis Centre<br>BLK 935 Tampines St 91. #01-333. S (520935)<br>Tel : 6789 8534 / Fax : 6784 5244           |
| <b>TM2</b>                 | <b>271 TAMPINES</b><br>NKF Dialysis Centre Supported by Wong Sui Ha Edna<br>BLK 271 Tampines St 21. #01-99. S (520271)<br>Tel : 6789 9878 / Fax : 6789 7336                             |
| <b>TPH</b>                 | <b>225 TOA PAYOH</b><br>NKF Dialysis Centre Supported by Toa Payoh Seu Teck Sean Tong<br>BLK 225 Toa Payoh Lorong 8. #01-54. S (310225)<br>Tel : 6254 2066 / Fax : 6251 9484            |
| <b>TP2</b>                 | <b>TOA PAYOH WEST CC</b><br>Seck Hong Choon-NKF Dialysis Centre<br>BLK 200 Toa Payoh Lorong 2. #03-01. S (319642)<br>Tel : 6970 4190 / Fax : 6970 4195                                  |
| <b>TWY</b>                 | <b>113 TECK WHYE</b><br>Leong Hwa Chan Si Temple-NKF Dialysis Centre<br>BLK 113 Teck Whye Lane. #01-666. S (680113)<br>Tel : 6769 0178 / Fax : 6769 9231                                |
| <b>UBK</b>                 | <b>19 UPPER BOON KENG</b><br>NKF Dialysis Centre Supported by Sakyadhita<br>Blk 19 Upper Boon Keng Rd #01-1220. S (380019)<br>Tel : 6743 1278 / Fax : 6743 1237                         |
| <b>URD</b>                 | <b>311 UBI</b><br>Foo Hai-NKF Dialysis Centre<br>BLK 311 Ubi Ave 1. #01-383 S (400311)<br>Tel : 6747 8864 / Fax : 6747 8823   |



Authorised Signature: \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_

|            |   |            |  |
|------------|---|------------|--|
| <b>HG1</b> | <b>114 HOUGANG</b><br>NKF Dialysis Centre Supported by Singapore Buddhist Welfare Services<br>BLK 114 Hougang Ave 1. #01-1298. S (530114)<br>Tel : 6382 6332 / Fax : 6383 0203        | <b>WCR</b> | <b>701 WEST COAST</b><br>The Hour Glass-NKF Dialysis Centre (West Coast Branch)<br>Blk 701 West Coast Rd #01-323. S (120701)<br>Tel : 6560 1184 / Fax : 6560 1076  |
| <b>HG2</b> | <b>628 HOUGANG</b><br>NKF Hougang-Punggol Dialysis Centre<br>BLK 628, Hougang Ave 8. #01-108. S (530628)<br>Tel : 6284 1877 / Fax : 6284 0867   | <b>WD1</b> | <b>Temporary Closure from 1st Jul 2025 onwards</b>   |
| <b>JE1</b> | <b>240C JURONG EAST</b><br>NKF Dialysis Centre Supported by Yuhua Grassroots Organisations<br>BLK 240C Jurong East Ave 1. #01-01. S (603240)<br>Tel : 6970 1847 / Fax : 6970 1849     | <b>WD2</b> | <b>365 WOODLANDS</b><br>NKFDialysis Centre Supported by The Singapore Contractors Association Ltd(SCAL)<br>BLK 365 Woodlands Ave 5. #01-490. S (730365)<br>Tel : 6362 4905 / 6362 3956 / Fax : 6362 5849 |
| <b>JW1</b> | <b>744 JURONG WEST</b><br>Sheng Hong Temple-NKF Dialysis Centre<br>BLK 744 Jurong West St 73. #01-19. S (640744)<br>Tel : 6794 1061 / Fax : 6794 1071                                 | <b>YS1</b> | <b>203 YISHUN</b><br>Toa Payoh Seu Teck Sean Tong-NKF Dialysis Centre<br>BLK 203 Yishun St 21. #01-239. S (760203)<br>Tel : 6759 4002 / Fax : 6759 4003  |
| <b>JW2</b> | <b>940 JURONG WEST</b><br>NKF Dialysis Centre Supported by The Srivadhanabhakdi Foundation<br>Blk 940 Jurong West Street 91. #01-441. S (640940)<br>Tel : 6316 6246 / Fax : 6316 6194 | <b>YS2</b> | <b>639 YISHUN</b><br>Le Champ-NKF Dialysis Centre (Yishun Branch)<br>BLK 639, Yishun St 61. #01-168, S (760639)<br>Tel: 6257 1860 / Fax: 6257 1650   |
| <b>KKT</b> | <b>NKF CENTRE</b><br>NKF Dialysis Centre Supported by Tay Choon Hye<br>81 Kim Keat Rd. S (328836)<br>Tel : 65062291 / Fax : 6254 9947   | <b>YS3</b> | <b>840 YISHUN</b><br>NKFDialysis Centre Supported by The Singapore Contractors Association Ltd(SCAL)<br>BLK 840 Yishun St 81. #01-382. S(760840)<br>Tel : 64813006 / Fax : 64813076                      |
| <b>KLA</b> | <b>43 BENDEMEER</b><br>NKF Dialysis Centre Supported by Kwan Im Thong Hood Cho Temple<br>BLK 43 Bendemeer Rd. #01-1018. S (330043)<br>Tel : 6291 3946 / Fax : 6291 3969               | <b>YS4</b> | <b>YISHUN COMMUNITY HOSPITAL</b><br>NKF Dialysis Centre Supported by KEPPEL<br>2 Yishun Central 2. #03-01 Yishun Community Hospital, S(768024)<br>Tel : 6970 4213 / Fax : 6970 4216                      |

**Accepted By:**

Authorised Signature: \_\_\_\_\_

Signatory's name : \_\_\_\_\_

Signatory's title : \_\_\_\_\_

Vendor's name : \_\_\_\_\_

Vendor's stamp : \_\_\_\_\_